

VISUAL APPARATUS CHECKS		TEST APPARATUS WRITTEN RECORD	
✓	<i>Check if ok</i>		
	FRAME/HARNESS ASSEMBLY	✓	<i>Check if ok</i>
	BREATHING BAG		LOW PRESSURE WARNING TEST 1.4 MBAR MAX.
	CYLINDER TEST		INHALATION VALVE TEST 10 MBAR +
	REGENERATIVE CANISTER		EXHALATION VALVE TEST 10 MBAR -
	SIMULATE ADDING DRAGERSORB		DRAIN VALVE TEST 10 MBAR
	RELIEF VALVE		LEAK TEST with POSITIVE PRESSURE 1 MBAR IN 60 SEC.
	COOLER		RELIEF VALVE TEST
	SIMULATE ADDING ICE		State Opening Pressure 2 TO 5 MBAR
	HOSES		HIGH PRESSURE LEAK TEST OPEN CYL. ICON
	COUPLING		CONSTANT METERING VALVE TEST
	FACE PIECE TEST		State Final Reading 1.5 TO 1.9 L/MIN
	SIMULATE DEFOGGING MASK		MINIMUM VALVE TEST
			State Opening Pressure 0.1 TO 2.5 MBAR
	CONNECTIONS CONNECTED OR GOOD		BYPASS VALVE TEST
	Cap on Drain Valve		ALTERNATE RELIEF VALVE TEST 2 TO 5 MBAR
	Drain Valve to Breathing Bag		LOW PRESSURE WARNING TEST
	Minimum Valve to Breathing Bag		State Warning Reading APPROX. 700 PSI
	Minimum Valve to O2 Supply Line		BATTERY TEST BATT FULL/ ICON
	Cylinder Connection		READY FOR USE
	Regenerative Canister Connections		DATE
	Relief Valve to Regenerative Canister		SERIAL NUMBER
	Crossover Hose Switchbox and Relief Valve		SIGNATURE
	Cooler to Breathing Bag		CYLINDER READING
	Distribution Hose		
	Breathing Hoses		
	Hose Adapter on RZ-25 or Test-It 6100		
			ZERO TESTER PRIOR TO TESTING
		VI	C
			VISUAL APPARATUS
			HOLE IN CORREGATED HOSE
			MASK PIN CUT
			MISSING SMALL RELIEF ORING
			DAMAGED DRAIN VALVE DISC
			FACE PIECE TOP BUCKLE UPSIDE DOWN (F/O)
			TEST APPARATUS
			HOLE IN CROSSOVER LINE
			MISSING 9V BATTERY
			HOLE BY-PASS LINE
			CORREGATED HOSE HOLE

